



Hangzhou Realy Tech Co., Ltd.

4th Floor, #12 Building, Eastern Medicine Town, Xiasha Economic & Technology Development, 310018 Hangzhou, Zhejiang, PEOPLE'S REPUBLIC OF CHINA

Tel: 86-571-5605 0793 Fax: 86-571-5605 0794

Certificate of Analysis

Product Name	Novel Coronavirus (SARS-Cov-2) Antigen Rapid Test Cassette (swab)		
Lot No.	202011004	Batch Quantity	800000 pcs
REF	K511416D	Expiry Date	2022-09

Test items	Test criteria	QC Result
Appearance	Aluminum foil bag and test device are intact and undamaged.	Conformity: 35/35
	Attachments should be complete.	Conformity: 35/35
Positive control (Low)	Conformity of positive control should be 3/3.	3/3
Positive control (Middle)	Conformity of positive control should be 3/3.	3/3
Negative control	Conformity of negative control should be 3/3.	3/3
Covid-Ag Buffer	Conformity of Covid-Ag Buffer should be 3/3.	3/3

Final QC Conclusion	This batch of product met the QC standards.
---------------------	---

QC Checked By: Hao Luang

Date: 2020.11.11

Approved By: Jian Yao

Date: 2020.11.11

